



CNEN/SP

ipen Instituto de Pesquisas
Energéticas e Nucleares

GOVERNO DO BRASIL

BETA AND GAMMA-RAY PERSONNEL MONITORING AT THE
CYCLOTRON OF THE INSTITUTO DE ENGENHARIA NUCLEAR
(IEN) BRAZIL

Patrícia W. FAJARDO, Letícia L. CAMPOS, Luiz Antonio R. da ROSA

IPEN-Pub-364

ABRIL/1992

SÃO PAULO

**BETA AND GAMMA-RAY PERSONNEL MONITORING AT THE CYCLOTRON
OF THE INSTITUTO DE ENGENHARIA NUCLEAR (IEN) BRAZIL**

Patrícia W. FAJARDO, Letícia L. CAMPOS, Luiz Antonio R. de ROSA

SERVIÇO DE CALIBRAÇÃO E DOSIMETRIA

**CNEN/SP
INSTITUTO DE PESQUISAS ENERGÉTICAS E NUCLEARES
SÃO PAULO – BRASIL**

Série PUBLICAÇÃO IPEN

INIS Categories and Descriptors

C55.00

BETA PARTICLES

GAMMA RADIATION

PERSONNEL DOSIMETRY

CYCLOTRONS

THERMOLUMINESCENT DOSEMETERS

CALIBRATION

IPEN-Doc-4234

Aprovado para publicação em 19/12/91

Note: A redação, ortografia, conceitos e revisão final são de responsabilidade do(s) autor(es).

**BETA AND GAMMA-RAY PERSONNEL MONITORING AT THE CYCLOTRON
OF THE INSTITUTO DE ENGENHARIA NUCLEAR (IEN) BRAZIL**

Patrícia W. FAJARDO¹, Letícia L. CAMPCS, Luiz Antonio R.
da ROSA²

COMISSÃO NACIONAL DE ENERGIA NUCLEAR-SP
INSTITUTO DE PESQUISAS ENERGÉTICAS E NUCLEARES
Caixa Postal 11049 - Pinheiros
05499 - São Paulo - BRASIL

ABSTRACT

Beta particle and gamma ray individual dose assessment was carried out at the cyclotron of the Instituto de Engenharia Nuclear (IEN), Brazil, using thermoluminescent dosimeters. The measurements were performed during the maintenance of activated components of the cyclotron and at the radioisotope production laboratory. All individual dose measurements were lower than the ICRP recommended limits for the skin and lens of the eyes.

(1) Instituto de Engenharia Nuclear - IEN

Rio de Janeiro, CEP 20001, CP 02186, Brasil

(2) Instituto de Radioproteção e Dosimetria - IRD

Rio de Janeiro, CEP 22780, CP 37750, Brasil

**MONITORAÇÃO PESSOAL BETA E GAMA NO CICLOTRON DO INSTITUTO
DE ENGENHARIA NUCLEAR (IEN) BRASIL**

Patrícia W. FAJARDO¹, Leticia L. CAMPOS, Luiz Antonio R.
da ROSA²

COMISSÃO NACIONAL DE ENERGIA NUCLEAR-SP
INSTITUTO DE PESQUISAS ENERGÉTICAS E NUCLEARES
Caixa Postal 11049 - Pinheiros
005499 - São Paulo - BRASIL

RESUMO

A determinação da dose individual devida a partículas beta e radiação gama foi executada no ciclotron do Instituto de Engenharia Nuclear (IEN), Brasil, usando dosímetros termoluminescentes. As medidas foram efetuadas durante a manutenção dos componentes ativados do ciclotron e no Laboratório de Produção de Radioisótopos. Todas as doses individuais determinadas estão abaixo dos limites recomendados pelo ICRP para a pele e o cristalino.

(1) Instituto de Engenharia Nuclear - IEN

Rio de Janeiro, CEP 20001, CP 02186, Brasil

(2) Instituto de Radioproteção e Dosimetria - IRD

Rio de Janeiro, CEP 22780, CP 37750, Brasil

1. INTRODUCTION

The Cyclotron of the Instituto de Engenharia Nuclear (IEN), Brazil, operates mainly for radioisotope production, as a neutron source and for irradiation damage analysis. The radiological risks associated with these activities result from exposure to γ rays, β particles and neutrons and due to contamination from induced radioactivity. Beta and γ ray exposure occurs during the maintenance of activated components of the cyclotron, due to ^{65}Zn produced by the activation of copper components. Other activities that involve the exposure to β and γ rays arise during radioisotope production, mainly from ^{67}Ga , which has ^{65}Zn as an undesirable subproduct. In this work, thermoluminescent (TL) dosimeters were used to assess β and γ ray individual doses received by IEN technicians, who were involved in the cyclotron maintenance and work at the radioisotope production laboratory. These dosimeters were specially developed for ^{137}Cs β and γ ray dose evaluation by Comissão Nacional de Energia Nuclear (CNEN), Brazil, during the radiological accident at Goiânia, Brazil ⁽¹⁾. Consequently, a new calibration and modifications to the measurement algorithm were necessary in order to determine the β and γ ray doses in the present circumstances.

2. DESCRIPTION OF THE DOSIMETER

Four $\text{CaSO}_4:\text{Dy}$.PTFE pellets 0.20 mm thick, produced at the Instituto de Pesquisas Energéticas e Nucleares (IPEN) ⁽²⁾, Brazil, are sealed in a black plastic film 16.5 $\text{mg}\cdot\text{cm}^{-2}$ thick and two pellets are shielded with PTFE filters 315 $\text{mg}\cdot\text{cm}^{-2}$ thick. These filters guarantee electronic equilibrium for ^{60}Co γ rays. The two remaining pellets

are unshielded. The shielded pellets are used for γ ray dose assessment, while the unshielded pellets are used to assess the β particle dose.

3. DOSIMETER CALIBRATION

3.1 Beta Particle Calibration

Three dosimeters were irradiated using the Beta Secondary Standard System of the Calibration Laboratory at IPEN, with ^{90}Sr - ^{90}Y (1.5 GBq), ^{204}Tl (5.1 MBq) and ^{147}Pm (81 MBq) sources. Compensation plastic filters provide field homogeneity within a diameter of 110mm at the calibration distances. The sources were calibrated in terms of absorbed dose rates, in air, at the Physikalisch Technische Bundesanstalt (PTB), Federal Republic of Germany, with a Primary Standard using an extrapolation chamber.

The energy response curve was determined for unshielded pellets (Fig. 1) and from this curve the calibration factor (absorbed dose in air/TL response) of the unshielded detector for the desirable β particle mean energy was evaluated. The relation between the TL responses of the shielded and unshielded pellets as a function of the mean energy of the β radiation was also determined and is presented in Fig. 2.

3.2 X-ray and Gamma ray Calibration

The X and γ ray radiation energy responses of the shielded and unshielded pellets were determined from 20 to 1250 keV and normalized to the ^{60}Co radiation response of the shielded pellets. X ray irradiation experimental conditions are given in Table 1. The energy

TABLE 1**Experimental Conditions for Irradiation by X-Rays.**

EFFECTIVE ENERGY (keV)	ADDED FILTRATION (mm)	HVL (mm)
25	2.0 Al	0.75 Al
30	2.0 Al	1.62 Al
34	2.0 Al	2.86 Al
45	4.0 Al	5.49 Al
55	0.2 Cu	0.35 Al
66	0.5 Cu	0.57 Al
70	0.5 Cu	0.7 Al
74	0.5 Cu	0.76 Al
96	1.0 Cu	1.42 Al
116	Th II	2.07 Cu
126	Th I	2.4 Cu
155	Th III	4.0 Cu




Fig. 1 - Unshielded Pellet Calibration Curve as a Function of the Beta Radiation Mean Energy.




Fig. 2 - Shielded Pellet Attenuation Factor as a Function of the Beta Radiation Mean Energy.

responses for shielded and unshielded pellets are in Fig. 3.

3.3 Algorithm for Evaluating Beta and Gamma ray Doses

The algorithm for evaluating the β and γ ray doses has the form

$$\begin{aligned} L_{\beta} + \alpha_1 L_{\gamma} &= L_{\mu} \\ \alpha_2 L_{\beta} + L_{\gamma} &= L_{s} \end{aligned} \quad (1)$$

where

- L_{β} is the response of the unshielded TL pellet for β particles (nC);
- L_{γ} is the response of the shielded TL pellet for γ rays (nC);
- α_1 is the ratio of the responses of the unshielded and shielded TL pellets for γ rays;
- α_2 is the ratio of the responses of the shielded and unshielded TL pellets for β particles;
- $\alpha_1 L_{\gamma}$ is the response of the unshielded TL pellet for γ rays (nC);
- $\alpha_2 L_{\beta}$ is the response of the shielded TL pellet for β particles (nC);
- L_{μ} is the total TL response of the unshielded pellet after irradiation with a β and γ ray source;
- L_s is the total TL response of the shielded pellet after irradiation with a β and γ ray source.

The α_1 energy dependence is the ratio of the energy response of the unshielded and shielded pellets for photons (Fig. 3). α_2 is obtained from Fig. 2.




Fig. 3 - Shielded and Unshielded Pellets as a Function of the X-Rays Effective Energy.

The β dose, D_β (Gy), and the γ ray dose, D_γ (Gy), may be evaluated using the relations

$$D_\beta = L_\beta \cdot f_{c\beta} \quad (2)$$

and

$$D_\gamma = L_\gamma \cdot f_{c\gamma} \quad (3)$$

where $f_{c\beta}$ and $f_{c\gamma}$ are, respectively, the β particle calibration factor ($\text{Gy} \cdot \text{nC}^{-1}$) of the unshielded pellets and the γ ray calibration factor ($\text{Gy} \cdot \text{nC}^{-1}$) of the shielded pellets. $f_{c\beta}$ is obtained from Fig. 1 and $f_{c\gamma}$ is the shielded pellet ^{60}Co γ ray calibration factor, $28.85 \times 10^{-5} \text{Gy} \cdot \text{nC}^{-1}$, corrected by the energy response curve of the shielded pellet (Fig. 3).

^{65}Zn decays emitting a β particle with a mean energy of 143 keV and a γ ray of 1115,5 keV. Therefore, from the curves of Fig. 2 and Fig. 3, the values of α_2 and α_1 may be obtained. The algorithm for this isotope has the form:

$$L_\beta + L_\gamma = L_\mu \quad (5)$$

$$0.0065 L_\beta + L_\gamma = L_g$$

This is the algorithm used for β and γ ray dose assessment during Cyclotron maintenance. $f_{c\beta}$ and $f_{c\gamma}$ are, respectively, $12 \times 10^{-4} \text{Gy} \cdot \text{nC}^{-1}$ and $28.85 \times 10^{-5} \text{Gy} \cdot \text{nC}^{-1}$.

When processing ^{67}Ga , the subproduct ^{65}Zn has also to be considered. ^{67}Ga emits only γ rays with several energies and intensities. Considering both ^{67}Ga and ^{65}Zn γ rays and ^{65}Zn β particles, the following algorithm was obtained:

$$L_{\beta} + 0.88 L_{\gamma} = L_{\mu} \quad (5)$$

$$0.0065 L_{\beta} + L_{\gamma} = L_{s}$$

$f_{c\gamma}$ and $f_{c\beta}$ are, respectively, $12 \times 10^{-4} \text{ Gy} \cdot \text{nC}^{-1}$ and $18.80 \times 10^{-5} \text{ Gy} \cdot \text{nC}^{-1}$.

4. MEASUREMENT PROCEDURES

The TL dosimeters were worn by IEN technicians, for 60 days, on the wrist and forehead. The wrist dosimeter was used to evaluate the dose to the hand and the forehead dosimeter for dose to the lens of the eyes. The average monthly doses for γ rays and β particles were evaluated for cyclotron maintenance and ^{67}Ga chemical processing, 1.85×10^9 Bq of ^{67}Ga was processed during the 60 days of monitoring.

5. RESULTS AND CONCLUSIONS

Table 2 presents the mean monthly β and γ radiation doses obtained during cyclotron maintenance. Table 3 presents the β and γ radiation monthly mean doses obtained during ^{67}Ga chemical processing. As observed, all values are lower than the ICRP⁽³⁾ maxima for the skin and for the lens of the eyes. These results illustrate the adequate sensitivity of the dosimeters which were used and the adequate radiation protection procedures during maintenance of the IEN cyclotron and ^{67}Ga processing.

TABLE 2

Beta and Gamma Radiation Month Mean Doses Evaluated During IEN Cyclotron Maintenance.

	Wristlet Dosimeter	Forehead Dosimeter
Gamma Dose (Gy)	2.29×10^{-4}	2.02×10^{-4}
Beta Dose (Gy)	5.75×10^{-4}	2.64×10^{-4}

TABLE 3

Beta and Gamma Radiation Month Mean Doses Evaluated During ^{67}Ga Chemical Processment.

	Wristlet Dosimeter	Forehead Dosimeter
Gamma Dose (Gy)	6.96×10^{-5}	5.26×10^{-5}
Beta Dose (Gy)	4.68×10^{-4}	1.08×10^{-4}

REFERENCES

1. CAMPOS, L.L. and DA ROSA, L.A.R. Cesium-137 Beta/Gama Discriminating Thermoluminescent Dosemeter. Radiat. Prot. Dosim, 25(1): 55, 1988 .
2. CAMPOS, L.L. and LIMA, M.F. Thermoluminescent CaSO_4Dy + Teflon Pellets for Beta Radiation Detection. Radiat. Prot. Dosim, 18 (2): 95, 1987 .
3. ICRP. Recommendations of the International Commission on Radiological Protection. Report 26. Oxford-Pergamon, 1977 .